

Air Ionizer Verification Record

Ionizer Verification Sequence Number: 08-045

WORKING STANDARD USED						
Asset/ISO #:	Manufacturer:	Model:	Serial No.	Calibration Date:	Calibration Due:	Calibration By:
25171	ION	775	6779	8/08/07	8/08/08	JPL

AIR IONIZER INFORMATION						
Asset/ISO #:	Manufacturer:	Model:	Serial No.	Verification Date:	Verification Due:	Verification By:
28516	MKS	5810i	10249	4-14-08	10-14-08	JPL 004
Inspector:	Location:	Owner:	Fail: Y/N ?	Cleaned: Y/N ?	Adjusted: Y/N ?	Prior Sequence#
BERJ ALLOJIAN	103/108A	Guy P.	N	N	N	N/A

VERIFICATION DATA						
HBM Sensitivity Level: <u>50</u> (from Table 1)						
Fan controller setting: <u>High</u> (High, Low, NA)						
Distance of ionizer from the charge plate: <u>24"</u>						
Ionizer Float Potential Tolerance \pm <u>50</u> Vdc. (from Table 1)						
Measured Float Potential values recorded below.						
1 <u>0</u> Vdc.	2 <u>0</u> Vdc.	3 <u>0</u> Vdc.	4 <u>0</u> Vdc.	5 <u>0</u> Vdc.	Comments:	
Ionizer Discharge Voltage Range: \pm 1000 Vdc to $< \pm$ <u>50</u> Vdc (from Table 1)						
Ionizer Discharge Time Tolerance: <u><20</u> seconds. (from Table 1)						
Measured Discharge Time in second(s) and recorded values below.						
1 (+1000 to +Vdc) <u>3.2</u> sec	2 (+1000 to +Vdc) <u>3.3</u> sec	3 (+1000 to +Vdc) <u>3.4</u> sec	4 (+1000 to +Vdc) <u>3.7</u> sec	5 (+1000 to +Vdc) <u>3.6</u> sec	Comments:	
1 (-1000 to -Vdc) <u>4.8</u> sec	2 (-1000 to -Vdc) <u>4.6</u> sec	3 (-1000 to -Vdc) <u>5.0</u> sec	4 (-1000 to -Vdc) <u>4.8</u> sec	5 (-1000 to -Vdc) <u>5.0</u> sec	Comments:	

Record any corrective action required to restored ionizer operation (cleaning, adjustment, replacement, etc.)

If Ionizer was replaced, indicate below the identification of replacement.

Asset/ISO #: _____ Manufacturer: _____ Model: _____ Serial No.: _____

Sequence number for verification of replacement Ionizer: _____

Record inspection schedule and rational for that schedule.